



Ref. Certif. No.

SI-7988IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product

Switch Mode Power Supply

Name and address of the applicant

PULS GmbH
Elektrastrasse 6, D-81925 München, Germany

Name and address of the manufacturer

PULS GmbH
Elektrastrasse 6, D-81925 München, Germany

Name and address of the factory

See page 2

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

 Additional Information on page 2Input: 100-240 Vac; 50-60 Hz; 1,4-0,6 A
Output: 24-28 Vdc; 5,0-4,3 A

Trademark / Brand (if any)

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

CP5.241-M1

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

A sample of the product was tested and found to be in conformity with

 Additional Information on page 2

IEC 60601-1:2005 + A1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0265/20 (2020-07-10)

This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services./

Date: 2020-07-10

Signature: Bojan Pečavar

Name and address of the factory:

1) PULS ELECTRONICS (SUZHOU) CO LTD
NO 1 RUI EN LANE XINGPU RD, SIP
SUZHOU, JIANGSU 215021,
CHINA

2) PULS INVESTICNI S R O
PRAZSKA 5639
430 01 CHOMUTOV,
CZECH REPUBLIC

Additional information (if necessary)

Date: 2020-07-10

Signature: Bojan Pečavar

